


28081T-1-3



ENVIRONMENTAL CONDITIONS

TEMP -40° TO +225°F
 PRESSURE
 VACUUM
 CHEMICALS
 OTHER

QC TESTING

ELECTRICAL HIPOT @ 150VAC
 VACUUM LEAK @ 1 X 10⁻⁸ STD cc He/sec
 PRESSURE LEAK @

CONSTRUCTION

INTERNAL 5294
 EXTERNAL
 POTTING 1469/1481

HOUSING SPECIFICATIONS

HSG P/N 6098-11-1
 THREAD 3/4"
 MAT'L BRASS
 TYPE NPT NIPPLE

LEAD WIRE / CONNECTOR DATA

CONTAINS:
 (12 PR) #24AWG TYPE 'T' TEFLON TC/, 10994-T-24-1.

- NOTES:
 1. NOT ALL WIRES SHOWN.
 2. MARK/TAG WIRES END TO END.

MASTER

UNLESS NOTED: All Dimension In Inches, Remove All Burrs, Break All Sharp Edges, Tolerances: .XX ±.01, .XXX ±.005

DOUGLAS ENGINEERING COMPANY
 14 Beach Street, Rockaway, NJ 07866
 PH: 201-627-8230 FX: 201-627-5798

ECN#	REV	APP'D	DATE	Title
				3/4" NPT NIPPLE FEEDTHRU
			Drawn TW Date 7/17/97 App'd <i>[Signature]</i> Date 7/18/97	Dwg No 28081T-1-3
				Rev. OR
Ref: XL SYSTEMS				SCALE: NTS SHEET 1 OF 1